

**Virginia Beach EMS  
LifePak 15 BLS Orientation**

**Pre-Test**

1. TRUE or FALSE

In Virginia Beach, the LP 15 powers up in the AED mode. It will automatically analyze the rhythm and advise that a shock is needed.

2. TRUE or FALSE

If you did NOT witness the arrest, you should perform 2 minutes of CPR prior to the first defibrillation.

3. TRUE or FALSE

It is not important to use the CPR Metronome during a cardiac arrest.

4. The AED has charged but it is unsafe to shock. How do you disarm the shock?

5. TRUE or FALSE

The LP15 AED shock sequence is 200j biphasic for the first shock, 300j biphasic for the second shock, and 360j biphasic for the third and subsequent shocks.

6. How do you enter manual mode?

7. You have entered a menu screen and want to get back to the default screen. What do you do?

8. TRUE or FALSE

The LifePak12 and LifePak 15 work with either Medtronic Quick-Combo pads or R2 pads.

9. How do you know which type of defib pads to use with the LifePak12, LifePak15 or MRL?

10. TRUE or FALSE

In Virginia Beach, you must take a manual BP prior to using the Non-invasive blood pressure (NiBP) feature of the LP15.